



Bid Notice Abstract

Request for Quotation (RFQ)

Reference Number 8977587
Procuring Entity PHILIPPINE RICE RESEARCH INSTITUTE
Title Procurement of One (1) Unit Sonicator (ALT22-04-0033T)
Area of Delivery Nueva Ecija

Solicitation Number:	PR no. 2022-08-1506	Status	Active
Trade Agreement:	Implementing Rules and Regulations		
Procurement Mode:	Negotiated Procurement - Small Value Procurement (Sec. 53.9)	Associated Components	1
Classification:	Goods	Bid Supplements	0
Category:	Laboratory Supplies and Equipment		
Approved Budget for the Contract:	PHP 200,000.00	Document Request List	6
Delivery Period:	30 Day/s		
Client Agency:		Date Published	02/09/2022
Contact Person:	Clarivel Ordonia Seguritan Admin Officer IV Maligaya Science City of Muñoz Nueva Ecija Philippines 3119 63-44-4560650 63-44-4560650 coseguritan@philrice.gov.ph	Last Updated / Time	02/09/2022 00:00 AM
		Closing Date / Time	09/09/2022 17:00 PM

Description

1 unit Sonicator
 25-L Capacity
 - Dimensions: 640 x 360 x 370 mm maximum
 - Capacity: 25 L
 - Controller type: Digital
 - Ultrasonic frequency: 53 KHz
 - Temperature range: ambient to 85 deg celsius
 - With Programmable timer
 - Timer range: 1-99 min
 - Operating environmental temperature: 20-35 deg celsius
 - Outer housing materials: Stainless steel or better
 - Interior tank material: Stainless steel
 - Cover: Stainless steel with handle
 - Ultrasonic power: 100W
 - Heating power: 275W
 - Degas power: 100W
 - With Soundproof chamber
 - Noise level: less than 60 decibels or lower (but will not compromise the ultrasonic power)
 - Power requirements: 220V/50-60 Hz
 - With drainage and valve (easy to close and open ball valve)
 - With compatible AVR/UPS
 - With one year warranty

Other Information

Email Address: frederickcapinpin22@gmail.com

Contact no.: 0907-925-5043

Created by Clarivel Ordonia Seguritan
Date Created 01/09/2022

The PhilGEPS team is not responsible for any typographical errors or misinformation presented in the system. PhilGEPS only displays information provided for by its clients, and any queries regarding the postings should be directed to the contact person/s of the concerned party.