



28 May 2024

Bid Bulletin No. 1

**Subject: SUPPLY AND DELIVERY OF GROWTH CHAMBER AND SEED INCUBATOR
PB 24-02-24**

This bid bulletin is being issued to clarify and amend Section VI. Schedule of Requirements and Section VII. Technical Specifications in the bidding document.

Please refer to the attached documents for complete details.

Section VII (Revised Technical Specifications) must be included in the bid proposal as part of the Eligibility and Technical Component (1st envelope) to signify conformity to the revisions.

Please be guided accordingly.

LEYLANI M. JULIANO

Vice Chairperson, BAC for Goods and Services

Better Rice Communities.

Email: prri.mail@mail.philrice.gov.ph **Text Center:** (+63) 917-111-7423

Websites: www.philrice.gov.ph, www.pinoyrice.com

Social Media: web.facebook.com/DAPhilRice

Liaison Office: 3rd Floor, ATI building, Elliptical Road, Diliman, Quezon City

Mobile No.: (+63)928-915-9628



SECTION VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item no.	Description	Unit	Qty	Total ABC	Division	Delivery Period
1	Growth Chamber	unit	2	7,200,000.00	CBC	120 calendar days upon receipt of PO/NTP
2	Seed Incubator	unit	1	70,000.00	CBC	60 calendar days upon receipt of PO/NTP

TOTAL ABC = P7,270,000.00

SECTION VII. Technical Specifications

Item	Specification	Statement of Compliance
		<i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i>
1	<p>Two (2) units Growth Chamber (with UPS and AVR, rewiring, wattage, installation, space)</p> <p>Plant Research Applications Tall Plant: Low Temp Plant-Derived Pharma Applications Tall Plant: Low Temp Form Reach-In Tiers 1 Tier External Area: 26.25 Ft2 External Dimensions: 105 In X 36 In X 89 In Growth Area: 1.5 Sqm Growth Height: 1650 Mm Temperature: 10 - 40 Degree Celsius (Light On) 4 - 40 Degree Celsius (Light Off)</p> <p>Programming: ADIAL Automatic Telephone Dialing System ARGUS Titan Control System AUTOW Auto Watering System AUX Auxiliary Circuits CMPLink Enables Full Argus Interaction SEN Temperature Sensor(s) UPS Controller UPS</p> <p>Construction: GH Growth Height Increases RECP Electrical Receptacles, Type & location SMC Split machine compartment</p> <p>Lighting: LMMCL Lamps Incorporate Dimming Ballasts</p> <p>Temperature: Low Temperature Operation</p>	

	<p>Ultra-Low Temperature Operation</p> <p>Refrigeration: ACSC Air-Cooled Self-Contained Condensing RAC Remote Outdoor Air-Cooled Condenser OACU Outdoor Air-Cooled Condensing Unit GLY Chilled Water/Glycol Heating/Cooling</p> <p>Humidity: ASNH Assisted Spray Nozzle Humidification USH Ultrasonic Humidification BDH Bypass Dehumidification</p> <p>Carbon Dioxide: CO2 Carbon Dioxide Additive Control SCRUB Carbon Dioxide Scrubber</p> <p>Inclusions:</p> <ol style="list-style-type: none">1. Compatible UPS with AVR anti-surge compatible to run the machine for at least 3-4 hrs.2. Spare lamp (3 unit compatible) <p>Warranty:</p> <ol style="list-style-type: none">1. Parts and Components: Two (2) years2. Services: Two (2) years <p>Other requirements:</p> <ol style="list-style-type: none">1. With calibration certificate upon delivery2. One-time calibration after one (1) year with Certificate3. The company must be at least 5 years in the business4. The brand must be at least 5 years in the market (supported by proof of sales)5. Certificate of training for personnel6. Onsite training on basic operations and troubleshooting7. ISO certification8. Certificate of availability of parts and components for 5 years	
2	<p>One (1) unit Seed Incubator</p> <p>Transparent box, with temperature and humidity control for table top (0-90°Celsius), (0-50%)</p> <p>Size Approximately: Height: 10 in Length:16.3 in Width: 9.5 in</p> <p>Display: LCD Electrical Requirements: 220V</p> <p>Warranty:</p>	

	<div>1. Parts and Components: Two (2) years</div> <div>2. Services: Two (2) years</div> <div>Other requirements:</div> <div>1. The company must be at least 5 years in the business</div> <div>2. The brand must be at least 5 years in the market (supported by proof of sales)</div> <div>3. Certificate of Training for Personnel</div> <div>4. Onsite training on basic operations and troubleshooting</div>	
	*****nothing follows*****	