



28 February 2025

Bid Bulletin No. 1

Subject: SUPPLY AND DELIVERY OF ONE (1) UNIT GEL ELECTROPHORESIS SYSTEM & ONE (1) UNIT THERMAL CYCLER PB 25-01-09

This bid bulletin is being issued to clarify and amend Section VI. Schedule of Requirements and Section VII. Technical Specifications in the bidding document. This must be included in the bid proposal as part of the Eligibility and Technical Component (1st envelope) to signify conformity to the revisions.

Please refer to the attached documents for complete details.

Please be guided accordingly.

-sgd-

LUIS ALEJANDRE I. TAMANI

Chairperson, BAC for Goods and Services

Section VI. Schedule of Requirements

Lot no.	Description	Unit	Qty	Total ABC	Division	Delivery Period
1	Vertical Rig Gel Electrophoresis System	unit	1	220,000.00	GRD	90 calendar days upon receipt of PO/NTP
2	PCR Machine, Thermal Cycler	unit	1	500,000.00	GRD	

Section VII. Technical Specifications

Lot	Specification	Statement of Compliance
		<i>[Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i>
1	<p>One (1) unit of Vertical Rig Gel Electrophoresis System</p> <p>Specifications:</p> <ul style="list-style-type: none">- Polyacrylamide Mini Gel Electrophoresis System- Screens up to 204 samples simultaneously- Gel sizes: 33cm (w) x 10cm (h), with 102 well microtiter-compatible comb to run 96 samples <p>Accessories:</p> <ul style="list-style-type: none">- Electrophoresis cell with cooling chamber and coolant recirculation ports- Electrodes- Silicone gaskets- Safety cover with power leads- White spring clamps- 2 combs- 2 sets of spacers- 2 sets of glass plates- 2 lengths gasket <p>Warranty:</p> <p>1. Parts and Components: Two (2) years</p>	

2	<p>One (1) Unit PCR Machine, Thermal Cycler</p> <p>Specifications:</p> <ul style="list-style-type: none">- Sample capacity: 1 x 96-well plate, 96 x 0.2 ml tubes, 0.2 ml tube strips- Temperature range: 4 to 100 C with accuracy of 0.5 °C per target- Heating and Cooling method: Peltier- Maximum ramp rate: 4 °C/sec- Average ramp rate: 2.5 °C/sec- Uniformity: 0.2°C-0.5°C- Gradient Capacity: Yes- Gradient Range: 30-100°C- Input power: 100-240VAC, 50-60Hz, 700 W- Display: VGA color touchscreen- User Interface: USB- Number of programs: Unlimited with USB flash drive <p>Other Requirements:</p> <ol style="list-style-type: none">1. Warranty: 2 years, after one (1) year Calibration/validation, and preventive maintenance2. With calibration / validation certificate or equivalent from manufacturer3. Onsite training on basic operation and troubleshooting4. The company must be at least five (5) years in the market. The brand must be at least ten (10) years in the market <p>Warranty:</p> <ol style="list-style-type: none">1. Parts and Components: Two (2) years2. Services: Two (2) years	
	<p>Important Notice:</p> <p>The supplier must provide evidence that the specifications are met through submission of brochure and complete technical specifications (SEPARATE DOCUMENT, use your company letterhead) indicating ALL the technical specifications of the product being offered, but not limited to Brand, Model, Place of Origin, Features, Inclusions & Warranty.</p> <p>NON-SUBMISSION of the complete technical specifications can lead to improper evaluation of your offer and rejection of bids. The bidder shall not be allowed to submit additional documents to correct any defects in the bid submitted.</p> <p>*****nothing follows*****</p>	

I hereby verify to comply with all the above requirements.

Signature over the printed name of the authorized representative

Company name

Date